

## Section E



# Siemens Replacement and Siemens Compatible Ion Chambers

26 February 2010

*-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-*

ADVANCED INSTRUMENT DEVELOPMENT, INC. 2545 CURTISS STREET DOWNERS GROVE, ILLINOIS 60515-4059  
TELEPHONE: (630) 271-9729 FACSIMILE: (630) 271-9995 [www.aidxray.com](http://www.aidxray.com)

# Section E



<b>MODEL:</b> ICX701	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61142A	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.22 IN (5.7mm)	
<b>APPLICATION:</b>	
REPLACES SIEMENS MODEL 81 67 538 X1651	

<b>MODEL:</b> ICX722	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61176N	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.28 IN (7.1mm)	
<b>APPLICATION:</b>	
CANON DIGITAL RADIOGRAPHIC REMOVABLE REMOTE PRE-AMP SIEMENS COMPATIBLE	

<b>MODEL:</b> ICX724	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61176C	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.22 IN (5.7mm)	
<b>APPLICATION:</b>	
FUJI FCR 5501 17X17 CHEST STAND AID COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section E



<b>MODEL:</b> ICX725L	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61176C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.22 IN (5.7mm)	
<b>APPLICATION:</b> FUJI FCR 5502 17X17 TABLE HEAD-END LEFT CONFIGURATION  SIEMENS COMPATIBLE	

<b>MODEL:</b> ICX725R	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61176C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.22 IN (5.7mm)	
<b>APPLICATION:</b> FUJI FCR 5502 17X17 TABLE HEAD-END RIGHT CONFIGURATION  SIEMENS COMPATIBLE	

<b>MODEL:</b> ICX726	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61176S	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.13 IN (3.3mm)	
<b>APPLICATION:</b> FUJI VELOCITY U 17 X17 CHEST STAND  SIEMENS COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section E



<b>MODEL:</b> ICX738	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61176C	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> IMIX DIGITAL RADIOGRAPHY SYSTEM  SIEMENS COMPATIBLE	

<b>MODEL:</b> ICX739	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61166C	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> IMAGING DYNAMICS CORP. DIGITAL CHEST STAND  SIEMENS COMPATIBLE	

<b>MODEL:</b> ICX740	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61176C	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.22 IN (5.7mm)	
<b>APPLICATION:</b> TRIXELL PIXIUM 4600 (USE INSTALLATION KIT 705172) AND TOSHIBA STB-01  REMOVABLE REMOTE PRE-AMP SIEMENS COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section E



<b>MODEL:</b> ICX753	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61176J	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> GENERIC L.H. LOAD SIDE -TURN POTS  SIEMENS COMPATIBLE	

<b>MODEL:</b> ICX754	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61176S	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.13 IN (3.3mm)	
<b>APPLICATION:</b> KONICA REGIUS 350 CR SYSTEM  SIEMENS COMPATIBLE	

<b>MODEL:</b> ICX942	
<b>FIELDS:</b> A=3, B=1, C=2	
<b>PRE-AMP:</b> 61176C	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> REPLACES SIEMENS MODEL 4133 518 X 1181	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section E



<b>MODEL:</b> ICX2701	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61166C	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.22 IN (5.7mm)	
<b>APPLICATION:</b> REPLACES SIEMENS MODEL 81 67 538 X1651  SIEMENS COMPATIBLE	

<b>MODEL:</b> ICX2722	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61176S	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.13 IN (3.3mm)	
<b>APPLICATION:</b> CANON DIGITAL RADIOGRAPHIC SYSTEM  SIEMENS COMPATIBLE	

<b>MODEL:</b> ICX3722	
<b>FIELDS:</b> A=2, B=1, C=3	
<b>PRE-AMP:</b> 61176S	
<b>OUTPUT:</b> SUMMED DC LEVEL	
<b>THICKNESS:</b> 0.13 IN (3.3mm)	
<b>APPLICATION:</b> CANON DIGITAL RADIOGRAPHIC SYSTEM  SIEMENS COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-