

# Section H



# Spot Ion Chambers

22 December 2009

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

ADVANCED INSTRUMENT DEVELOPMENT, INC. 2545 CURTISS STREET DOWNERS GROVE, ILLINOIS 60515-4059  
TELEPHONE: (630) 271-9729 FACSIMILE: (630) 271-9995 [www.aidxray.com](http://www.aidxray.com)

# Section H



<b>MODEL:</b> ICX100	
<b>FIELDS:</b> SPOT	
<b>PRE-AMP:</b> 61125A	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.34 IN (8.6mm)	
<b>APPLICATION:</b> REPLACES GE MODEL 14 AND MODEL 63 SPOT FILM DEVICE  AID COMPATIBLE	

<b>MODEL:</b> ICX101	
<b>FIELDS:</b> SPOT	
<b>PRE-AMP:</b> 61125A	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.34 IN (8.6mm)	
<b>APPLICATION:</b> REPLACES GE MODEL 88 SPOT FILM DEVICE  AID COMPATIBLE	

<b>MODEL:</b> ICX158	
<b>FIELDS:</b> SPOT 3 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61125C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> CONTINENTAL R/F TABLE TOP OF PRE-AMP COVER FLUSH WITH FRAME AID COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX161	
<b>FIELDS:</b> SPOT 2.0 IN (51mm) SQ.	
<b>PRE-AMP:</b> 61125C	
<b>OUTPUT:</b> NEGATIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> BENNETT SPOT FILM DEVICES  TREX-BENNETT COMPATIBLE	

<b>MODEL:</b> ICX164	
<b>FIELDS:</b> SPOT 3.5 IN (89mm) DIA.	
<b>PRE-AMP:</b> 61059F	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.34 IN (8.6mm)	
<b>APPLICATION:</b> AMERISYS, ELSCINT, XONICS SPOT FILM DEVICE  AID COMPATIBLE	

<b>MODEL:</b> ICX165	
<b>FIELDS:</b> F1 IS ALWAYS ON F1 = 1.8 IN (46mm) X 1.5 IN (38mm) F2 = SPOT 3.5 IN (89mm) DIA.	
<b>PRE-AMP:</b> 61059C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.34 IN (8.6mm)	
<b>APPLICATION:</b> AMERISYS, ELSCINT, XONICS DUAL-FIELD SPOT  AID COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX166	
<b>FIELDS:</b> A=1, B=2, C=3	
<b>PRE-AMP:</b> 61135A	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> BOTTOM OF PRE-AMP FLUSH WITH FRAME	
AID COMPATIBLE	

<b>MODEL:</b> ICX172	
<b>FIELDS:</b> SPOT 3.0 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61125B	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.34 IN (8.6mm)	
<b>APPLICATION:</b> COMMON SMALL SPOT FILM DEVICES	
AID COMPATIBLE	

<b>MODEL:</b> ICX173	
<b>FIELDS:</b> SPOT OFFCENTER 3.0 IN SQ.	
<b>PRE-AMP:</b> 61125B	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> COMMON SPOT FILM DEVICES	
AID COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX178	
<b>FIELDS:</b> SPOT 3.0 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61125B	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.34 IN (8.6mm)	
<b>APPLICATION:</b> EUREKA SPOT FILM DEVICE	
AID COMPATIBLE	

<b>MODEL:</b> ICX179	
<b>FIELDS:</b> SPOT 3.0 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61059F	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.34 IN (8.6mm)	
<b>APPLICATION:</b> MEDSTONE INTERNATIONAL	
AID COMPATIBLE	

<b>MODEL:</b> ICX186	
<b>FIELDS:</b> SPOT 2.0 IN (51mm) SQ.	
<b>PRE-AMP:</b> 61125B	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> OLDELFT SIMULIX	
AID COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX188	
<b>FIELDS:</b> SPOT 2.25 IN	
<b>PRE-AMP:</b> 61125B	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> SIEMENS SPOT FILM  AID COMPATIBLE	

<b>MODEL:</b> ICX194	
<b>FIELDS:</b> SPOT 3.0 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61125B	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.34 IN (8.6mm)	
<b>APPLICATION:</b> ACOMA SPOT FILM  AID COMPATIBLE	

<b>MODEL:</b> ICX195	
<b>FIELDS:</b> A-B-B+C=2-1-3	
<b>PRE-AMP:</b> 61112A	
<b>OUTPUT:</b> NEGATIVE RAMP	
<b>THICKNESS:</b> 0.28 IN (7.1mm)	
<b>APPLICATION:</b> PICKER SPOT FILMER	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX199	
<b>FIELDS:</b> SPOT 3.0 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61125C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> BOTTOM OF PRE-AMP CHASSIS FLUSH WITH FRAME	
AID COMPATIBLE	

<b>MODEL:</b> ICX200	
<b>FIELDS:</b> SPOT 3.0 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61125A	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> REPLACES PMT IN GE SPOT FILM DEVICE MODEL 14 AND MODEL 63	

<b>MODEL:</b> ICX201	
<b>FIELDS:</b> SPOT 3.0 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61125A	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.34 IN (8.6mm)	
<b>APPLICATION:</b> REPLACES GE MODEL 88 SPOT	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX202	
<b>FIELDS:</b> SPOT 3.0 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61125A	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> REPLACES PMT IN GE SATELLA X SPOT FILM DEVICE	

<b>MODEL:</b> ICX215	
<b>FIELDS:</b> A=3, B=2, C=1	
<b>PRE-AMP:</b> 61156A	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> MODEL 88/85 SPOT FILM DEVICE REPLACES GE MODEL NOS. 2224773, 2225330, 2331090, 2331095, 46-250767G1/G2, 46-303850G1, 46-316719G1, 46-316729G1	

<b>MODEL:</b> ICX258	
<b>FIELDS:</b> SPOT 3.0 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61125C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> GENERIC SPOT FILM DEVICE TOP OF PRE-AMP COVER FLUSH WITH FRAME  GE COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX286	
<b>FIELDS:</b> SPOT 2.0 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61125B	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> OLDELFT SIMULIX/  GE COMPATIBLE	

<b>MODEL:</b> ICX299	
<b>FIELDS:</b> SPOT 3.0 IN (76mm) SQ.	
<b>PRE-AMP:</b> 61125C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> BOTTOM OF PRE-AMP FLUSH WITH FRAME  GE COMPATIBLE	

<b>MODEL:</b> ICX301	
<b>FIELDS:</b> SPOT (2 IN SQ.)	
<b>PRE-AMP:</b> 61084E	
<b>OUTPUT:</b> POSITIVE RAMP (RANGE: +2VDC TO +45VDC)	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> EXIT-TYPE PEDIATRIC / SPOT FILM DEVICE  PHILIPS COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX310	
<b>FIELDS:</b> 3-2-1=A-(B+C)-C	
<b>PRE-AMP:</b> 61084C	
<b>OUTPUT:</b> POSITIVE RAMP (RANGE: +2VDC TO +45VDC)	
<b>THICKNESS:</b> 0.34 IN (8.6mm)	
<b>APPLICATION:</b> SPOT REPLACEMENT PHILIPS CONNECTOR REPLACES PHILIPS MODEL 9803 509 30002	

<b>MODEL:</b> ICX661	
<b>FIELDS:</b> SPOT 2.0 IN (51mm) SQ.	
<b>PRE-AMP:</b> 61125C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> LARGE SPOT FILM DEVICES  GTR LABS (AID) COMPATIBLE	

<b>MODEL:</b> ICX801	
<b>FIELDS:</b> SPOT (2 IN SQ.)	
<b>PRE-AMP:</b> 61125B	
<b>OUTPUT:</b> NEGATIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> EXIT-TYPE PEDIATRIC / SPOT FILM DEVICE  PICKER COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX907	
<b>FIELDS:</b> SPOT 3.0 IN (76mm)	
<b>PRE-AMP:</b> 61125B	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b>	
AID COMPATIBLE	

<b>MODEL:</b> ICX920	
<b>FIELDS:</b> SPOT 1.57 IN (40mm) DIA	
<b>PRE-AMP:</b> 61059G	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b>	
EXIT-TYPE R.P. KINCHLOE SKULL CHAMBER	
PHILIPS COMPATIBLE	

<b>MODEL:</b> ICX929	
<b>FIELDS:</b> SPOT 2.5-INCH DIA.	
<b>PRE-AMP:</b> 61125B	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b>	
ROUND SPOT	
REMOTE PRE-AMP AID COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX930	
<b>FIELDS:</b> SPOT 3-IN SQ.	
<b>PRE-AMP:</b> 61125B	
<b>OUTPUT:</b> NEGATIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b>  PICKER 1717 SPOT FILM DEVICE	

<b>MODEL:</b> ICX931	
<b>FIELDS:</b> 3-FLD SPLIT CENTER	
<b>PRE-AMP:</b> 61135F	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> 3-FIELD WITH SPLIT CENTER FOR FORMATTED SPOT FILM  AID COMPATIBLE	

<b>MODEL:</b> ICX938	
<b>FIELDS:</b> 3-FLD SPLIT CENTER	
<b>PRE-AMP:</b> 61135K	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> 3-FIELD WITH SPLIT CENTER FOR FORMATTED SPOT FILM  REMOVABLE REMOTE PRE-AMP NON-AID COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX972	
<b>FIELDS:</b> SPOT 3.0 IN (76mm)	
<b>PRE-AMP:</b> 61125C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> REPLACEMENT FOR SIEMENS EXPLORATOR 35 SFD WITH REMOVABLE REMOTE PRE-AMP  AID COMPATIBLE	

<b>MODEL:</b> ICX974	
<b>FIELDS:</b> SPOT 3.0 IN (76mm)	
<b>PRE-AMP:</b> 61125C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.34 IN (8.6mm)	
<b>APPLICATION:</b> REPLACEMENT FOR SIEMENS EXPLORATOR 25 SFD WITH REMOVABLE REMOTE PRE-AMP  AID COMPATIBLE	

<b>MODEL:</b> ICX979	
<b>FIELDS:</b> SPOT 3.0 IN (76mm)	
<b>PRE-AMP:</b> 61125C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.28 IN (7.1mm)	
<b>APPLICATION:</b> REPLACEMENT FOR VACUTEC BAK 70 142  AID COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-

# Section H



<b>MODEL:</b> ICX1106	
<b>FIELDS:</b> A=2, B=ACTIVE, C=3	
<b>PRE-AMP:</b> 61154N	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.13 IN (3.3mm)	
<b>APPLICATION:</b> 3-FIELD SPOT 374 X 374  AID COMPATIBLE	

<b>MODEL:</b> ICX1994	
<b>FIELDS:</b> Spot	
<b>PRE-AMP:</b> 61154D	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> VARIAN 2020 FPD FLUORO-SPOT ION CHAMBER  AID COMPATIBLE	

<b>MODEL:</b> ICX3958	
<b>FIELDS:</b> SPOT	
<b>PRE-AMP:</b> 61125C	
<b>OUTPUT:</b> POSITIVE RAMP	
<b>THICKNESS:</b> 0.4 IN (10.2mm)	
<b>APPLICATION:</b> NON-MAGNETIC SPOT FILM DEVICE  AID COMPATIBLE	

-SERVING THE MEDICAL X-RAY FIELD SINCE 1969-